



P/N ALPC-G14026-100



P/N ALCV-G14033-050

Attention: PICCOLO® Operator

Abaxis recommends the following Quality Control Schedule:

CLIA Waived Testers - must run two liquid controls per panel every lot change or at least once per calendar month, whichever comes first
(Run one vial ea L1+L2 of P/N ALPC-G14026-100)

CLIA Moderate Complexity Testers - are required to run two liquid controls at least once per calendar month
(Run one vial ea L1+L2 of P/N ALPC-G14026-100)
- are required to run a linearity check once every six months
(Run one vial ea S1+S2+S3 of P/N ALCV-G14033-050)

Please call Abaxis Technical Support at 1-800-822-2947 if you have any questions.

Quality Control materials can be ordered two ways directly from Nova-One Diagnostics

(1) By entering your order on line at NOVA-ONE.net (.net – not .com) clicking on the flashing red “Order Now” and filling out the on line order form; NOVA-ONE will ship and bill you or charge your credit card whichever you prefer.....**(Remember to finish the form at the bottom 3+5= 8 to make sure your order goes thru)**

OR

(2) By simply writing your order on a blank clinic letterhead. The customer faxes it to NOVA –ONE (Fax# 818-348-9696). NOVA-ONE requests the clinics use their letterhead stationary to ensure NOVA-ONE has a legible “ship to” address and phone.

Catalog No.

Products

ALPC-G14026-100

NOVA-ONE Diagnostics Chemistry Controls Level 1 – (6) Vials x 1mL, Level 2 – (6) Vials x 1mL
(Includes all analytes in each sample vial)

ALCV-G14033-050

NOVA-ONE Diagnostics Verification Samples Kit Set of 3 Samples x 3 Vials each x 0.5mL
(Includes all analytes in each sample vial)