

Product Number ALCV-G14033-050  
Lot # 0808024 (Contents; Samples: 0808024-1, -2, -3)  
Expiration Date: 09 2009

3 Levels x 3 vials x 0.5mL

**Intended Use**

**NOD®** Verification Samples are human Liquid assayed serum samples used to accomplish analyzer validation according to CLIA guidelines. The laboratory testing procedures are listed in the following text.

**Summary and Principles**

The use of independent quality verification sample materials is indicated as an objective assessment of the precision of methods and techniques in use and is an integral part of good laboratory practices. Three levels of verification samples are available to allow performance monitoring within the clinical range.

**Reagent Composition**

This product is prepared from purified human serum to which biochemical material (human and animal tissue extracts); drugs, chemicals, stabilizers and preservatives have been added. The verification samples are in a prepackaged liquid form to avoid potential error or contaminate being introduced during reconstitution.

**Storage and Stability** To achieve maximum shelf life for the kit store unopened at ≤ -20°C. Store vials away from the light.

**Thawed and Unopened:** The Verification Samples can be used for up to **14 days** when stored unopened 2-8°C. For optimum Bilirubin and CO<sub>2</sub> stability avoid prolonged exposure of the Verification Sample vials to ambient air / room temperature / light.

**Procedure**

The verification samples should be treated the same as a patient sample and run according to the instructions accompanying the instrument, kit, or reagent being used. Before sampling the verification samples should be mixed thoroughly but gently.

Thaw **NOD®** Verification Samples at room temperature (18-25° C) for 30 minutes or until completely thawed. Mix the vial thoroughly by inverting several times, before sampling gently swirls until homogeneous with no visible signs of precipitate. Avoid vigorous shaking. After sampling the Verification samples should be promptly re-capped and stored a 2-8°C. Dispose of at the end of day or upon completion of data collection.

**Limitation of Procedure**

(a) This product should not be used past the expiration date (b) if there is evidence of microbial contamination in the verification samples or excessive turbidity discard the vial (c) This Product is not intended for use as a standard.

The assay values recovered in the laboratory are method dependent and reflect reagent, method and technique and instrument variations. If methods and / or reagents are changed or modified the resulting assay value may be different.

**Assignment of Values**

The mean values printed on the circular were derived from replicate analyses on the Piccolo Clinical Chemistry Analyzer and are specific for this lot of Liquid Assayed Chemistry Verification Samples.

**Specific Performance Characteristics**

To ensure the reliability and usefulness of the verification samples, the product must be properly handled and stored as described.

Individual donor units used in the preparation of this product have been tested by FDA approved methods for anti-HIV 1 & 2, HBsAg, anti-HCV, HIV-1 antigens and Syphilis and found non-reactive. No test method can offer complete assurance that products derived from human source material will not transmit infectious diseases. Therefore, this product should be considered potentially infective and be treated in the same manner as a patient specimen.

**Ordering Information:** Verification P/N ALCV-G14033-050 or Control P/N ALPC-G14026-100  
By Ordering On Line at **NOVA-ONE.NET**; or Fax **818-348-9696**  
**NOVA-ONE Diagnostics 4987 Campo Road Woodland Hills, California 91364**

**Assigned Mean Values** (Representative Values)

**Lot # 0808024 (Contents; Samples: 0808024-1, - 2, - 3)**  
 Expiration Date: 09 2009

<b>METHOD: Abaxis Piccolo</b>	<b>Sample 1</b>	<b>Sample 2</b>	<b>Sample 3</b>
<b>Analyte, Units</b>	<b>Mean</b>	<b>Mean</b>	<b>Mean</b>
ALT/SGPT U/L	21	752	1475
Albumin g/dL	1.60	4.11	6.08
Alkaline Phosphatase (ALP) U/L	25	1072	1980
Amylase, Pancreatic U/L	25	1224	2344
Aspartate Aminotransferase (AST/SGOT) U/L	27	801	1545
Bilirubin – Direct mg/dL	0.25	2.46	4.6
Bilirubin – Total mg/dL	0.4	2.9	5.4
BUN (Urea Nitrogen) mg/dL	5	69	148
Calcium Total mg/dL	4.5	10.4	14.7
Carbon Dioxide (CO <sub>2</sub> ) mEq/L	34	22	7
Chloride mEq/L	82	107	130
HDL Cholesterol mg/dL	20	51	82
Total Cholesterol mg/dL	56	222	374
Creatine Kinase (CK) U/L	40	1864	3642
Creatinine mg/dL	0.6	8.7	16.8
GGT U/L	18	1267	2362
Glucose mg/dL	27	337	649
Lactate Dehydrogenase (LDH) U/L	73	409	811
Magnesium mg/dL	0.6	3.5	7.2
Phosphorus mg/dL	15.4	8.2	1.1
Potassium mEq/L	2.0	4.8	7.3
Protein, Total g/dL	2.8	6.9	10.9
Sodium mEq/L	117	141	165
Triglycerides mg/dL	41	197	354
Uric Acid mg/dL	0.9	7.1	14.4

**Ordering Information:** Verification P/N ALCV-G14033-050 or Control P/N ALPC-G14026-100  
 By Ordering On Line at **NOVA-ONE.NET**; or Fax **818-348-9696**  
**NOVA-ONE Diagnostics 4987 Campo Road Woodland Hills, California 91364**